

آزمایشگاه کولایف تهران

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.32	
Mean	267.74	Mean	488.74
SD	7.61	SD	6.44
CV	2.84	CV	1.32
Day-Run	Value	Day-Run	Value
1 - 1	275	1 - 1	481
2 - 1	260	2 - 1	490
3 - 1	258	3 - 1	494
4 - 1	265	4 - 1	491
6 - 1	258	6 - 1	480
8 - 1	267	8 - 1	494
9 - 1	280	9 - 1	502
11 - 1	269	11 - 1	495
13 - 1	283	13 - 1	491
14 - 1	277	14 - 1	493
15 - 1	260	15 - 1	480
16 - 1	265	16 - 1	491
20 - 1	270	20 - 1	495
21 - 1	278	21 - 1	493
22 - 1	267	22 - 1	485
23 - 1	265	23 - 1	485
24 - 1	265	24 - 1	480
25 - 1	265	25 - 1	483
27 - 1	260	27 - 1	483

Westgard	تفسیر	ALP	آزمایش	mindray BS-800M	دستگاه	9	ماه	1400	سال
Normal Control	Randox	Mean	270	SD	15.3	Exp. Date	1400/5/21	Lot. No	1489UN

