

آزمایشگاه پاتوبیولوژی و ژنتیک فرجاد

Control Chart

Date-Run	Value
1399/5/1-1	153
1399/5/2-1	150
1399/5/8-1	150
1399/5/8-2	147
1399/5/21-1	148
1399/5/21-2	155
1399/5/21-3	142
1399/5/21-4	159
1399/5/21-5	159
1399/5/21-6	153
1399/5/21-7	158
1399/5/27-1	148
1399/5/28-1	152
1399/6/1-2	150
1399/6/6-1	149
1399/6/11-1	152
1399/6/22-1	145
1399/6/25-1	140
1399/6/31-1	148
1399/7/8-3	150
1399/7/14-1	145
1399/7/15-1	146
1399/7/19-1	142
1399/8/1-3	152
1399/8/10-1	154
1399/8/11-2	147
1399/8/12-1	146
1399/8/15-2	161
1399/8/17-1	143
1399/8/18-1	159
1399/8/19-2	161
1399/8/20-1	156
1399/8/21-8	152
1399/8/22-2	153
1399/8/24-1	165

High					
mindray BS-800M	دستگاه	AST-ST	تست	1400/1/17	تاریخ
1:3.5s	قوانین تفسیر	Control	Delta D.P. Pathol-SF	Lot. No	674

Mean	151.96	SD	3.94	Calculated :	Mean	152.14	SD	4.57	CV %	3.00	Sigma	6.65	N	118
------	--------	----	------	--------------	------	--------	----	------	------	------	-------	------	---	-----



Date-Run	Value
1399/8/27-2	153
1399/8/28-2	149
1399/8/29-1	159
1399/9/1-4	155
1399/9/3-1	152
1399/9/4-1	151
1399/9/5-1	151
1399/9/6-2	149
1399/9/8-4	150
1399/9/9-1	152
1399/9/10-2	153
1399/9/11-3	146
1399/9/12-2	147
1399/9/15-3	156
1399/9/16-1	155
1399/9/17-2	152
1399/9/18-2	157
1399/9/19-3	155
1399/9/23-1	151
1399/9/24-2	148
1399/9/26-1	152
1399/9/27-3	152
1399/10/2-2	158
1399/10/8-5	157
1399/10/13-1	151
1399/10/21-9	154
1399/10/23-2	150
1399/10/24-3	147
1399/10/29-2	144
1399/10/30-1	154
1399/11/1-5	149
1399/11/4-2	145
1399/11/11-4	153
1399/11/12-3	156
1399/11/13-2	147
1399/12/4-3	149
1399/12/13-3	151.6
1399/12/16-2	151.6
1399/12/17-3	146
1399/12/18-3	150
1399/12/23-3	157
1399/12/25-2	157
1399/12/26-2	146
1399/12/27-4	156
1400/1/14-2	152.5

Date-Run	Value
1400/1/15-4	151.5
1400/1/16-3	152.5
1400/1/17-4	153
1400/1/18-4	149.6
1400/1/22-3	151.8
1400/1/28-3	155
1400/2/2-1	154.1
1400/2/6-1	149.9
1400/2/7-1	149
1400/2/7-1	149
1400/2/8-1	153
1400/2/9-1	154.8
1400/2/11-1	158
1400/2/13-1	144
1400/2/15-1	160
1400/2/18-1	159.7
1400/2/19-1	154.2
1400/2/20-1	153
1400/2/21-1	155.4
1400/2/21-2	155.4
1400/2/22-1	153
1400/2/26-1	151
1400/2/27-2	151.7
1400/2/29-1	157
1400/3/3-1	155.7
1400/3/6-1	148.6
1400/3/8-1	155.6
1400/3/9-1	157.8
1400/3/10-1	151.6
1400/3/11-1	155.8
1400/3/12-1	157
1400/3/13-1	160
1400/3/19-1	155.4
1400/3/22-1	155
1400/3/24-1	154.7
1400/3/26-1	149.6
1400/3/29-1	153.6
1400/3/30-1	150