

# آزمایشگاه پاتوبیولوژی و ژنتیک فرجاد

## Control Serum - Levey Jennings Chart

Normal		High		Low	
Mean	29.04	Mean	59.88	Mean	21.22
SD	1.33	SD	2.92	SD	0.55
CV	4.59	CV	4.87	CV	2.57
Day-Run	Value	Day-Run	Value	Day-Run	Value
1 - 1	29.61	1 - 1	54.43	1 - 1	20.28
2 - 1	28.13	2 - 1	57.88	2 - 1	21.03
3 - 1	28.94	3 - 1	53.69	3 - 1	21.6
4 - 1	27.17	4 - 1	60.75	4 - 1	20.62
6 - 1	27.48	6 - 1	58.87	6 - 1	21.17
8 - 1	28.96	8 - 1	62.43	8 - 1	20.96
9 - 1	26.94	9 - 1	60.03	9 - 1	21.5
10 - 1	31.41	10 - 1	60.91	10 - 1	21.83
12 - 1	31.41	12 - 1	64.18	12 - 1	21.44
13 - 1	29.79	13 - 1	64.2	13 - 1	22.49
13 - 2	27.76	13 - 2	57.7	13 - 2	21.29
17 - 1	28.27	17 - 1	57.7	17 - 1	21.72
18 - 1	28.95	18 - 1	61.52	18 - 1	21.42
19 - 1	28.79	19 - 1	59.99	19 - 1	20.44
20 - 1	29.88	20 - 1	61.11	20 - 1	20.84
30 - 1	30.08	30 - 1	61.26	30 - 1	20.88
31 - 1	30.08	31 - 1	61.26	31 - 1	21.17

Westgard	تفسیر	Anti tpo		آزمایش	Architect		دستگاه	3	ماه	1400	سال
Normal Control	multichem	Mean	28.52	SD	1.15	Exp. Date	1400/3/9	Lot. No	33402200		
High Control	multichem	Mean	59.79	SD	5.67	Exp. Date	1400/3/9	Lot. No	33102200		
Low Control	multichem	Mean	21.78	SD	1.08	Exp. Date	1400/3/9	Lot. No	33102200		

