

آزمایشگاه کولایف تهران

Control Serum - Levey Jennings Chart

Normal		High - CV : 1.66	
Mean	1.07	Mean	1.73
SD	0.02	SD	0.03
CV	1.84	CV	1.66
Day-Run	Value	Day-Run	Value
3 - 1	1.07	3 - 1	1.7
4 - 1	1.06	4 - 1	1.7
8 - 1	1.08	8 - 1	1.7
10 - 1	1.08	10 - 1	1.7
13 - 1	1.1	13 - 1	1.73
14 - 1	1.08	14 - 1	1.7
15 - 1	1.1	15 - 1	1.77
16 - 1	1.04	16 - 1	1.72
20 - 1	1.04	20 - 1	1.74
21 - 1	1.06	21 - 1	1.72
22 - 1	1.05	22 - 1	1.72
23 - 1	1.06	23 - 1	1.77
24 - 1	1.09	24 - 1	1.76
25 - 1	1.08	25 - 1	1.78
27 - 1	1.09	27 - 1	1.73

Westgard	تفسیر	Bili-D	آزمایش	mindray BS-800M	دستگاه	9	ماه	1400	سال
Normal Control	Randox	Mean	1.08	SD	0.08	Exp. Date	1400/5/21	Lot. No	1489UN

