

# آزمایشگاه کولایف تهران

## Control Serum - Levey Jennings Chart

Normal		High - CV : 1.38	
Mean	1.59	Mean	4.69
SD	0.03	SD	0.06
CV	1.95	CV	1.38
Day-Run	Value	Day-Run	Value
3 - 1	1.54	3 - 1	4.6
4 - 1	1.59	4 - 1	4.61
8 - 1	1.6	8 - 1	4.8
10 - 1	1.6	10 - 1	4.7
14 - 1	1.58	13 - 1	4.66
15 - 1	1.6	14 - 1	4.6
16 - 1	1.59	16 - 1	4.78
20 - 1	1.53	20 - 1	4.7
21 - 1	1.58	21 - 1	4.63
23 - 1	1.58	23 - 1	4.74
24 - 1	1.66	24 - 1	4.7
25 - 1	1.58	25 - 1	4.7
27 - 1	1.58	27 - 1	4.7

Westgard	تفسیر	Bili-T	آزمایش	mindray BS-800M	دستگاه	9	ماه	1400	سال
Normal Control	Randox	Mean	1.6	SD	0.1	Exp. Date	1400/5/21	Lot. No	1489UN

