


آزمایشگاه کولایف تهران

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
4 - 14	163.2	
4 - 15	167.3	
4 - 17	165.2	
4 - 19	161.2	
4 - 20	168.4	
4 - 21	159.4	
4 - 24	160.3	
4 - 26	166.2	
4 - 28	163.2	
4 - 29	160.2	10X
5 - 4	137.5	
5 - 5	136.5	
5 - 9	142.8	
5 - 11	152.1	
5 - 12	157.5	
14 - 1	147.1	
16 - 1	153.1	
17 - 1	165.2	
19 - 1	168.7	
21 - 1	162.9	10X
23 - 1	165.4	
24 - 1	165.6	
25 - 1	156.4	
26 - 1	151.4	
30 - 1	162.8	
31 - 1	162.8	
1 - 1	163.9	
2 - 1	160	
4 - 30	161.9	
4 - 31	160.9	
6 - 1	177.6	
7 - 1	166.1	
8 - 1	144.5	
11 - 1	161.3	
13 - 1	167.3	

Westgard	تفسیر	Free BHCG	آزمایش	Cobas Immunoanalyzer	دستگاه	6	ماه	1400	سال
Control	Randox Maternal	Mean	158.9	SD	15	Exp. Date	1399/12/24	Lot. No	7703MS
Next Month	Mean	161.33	SD	10.56	CV	6.54			



Day-Run	Value	Description
14 - 2	180.1	
15 - 1	184.8	
16 - 2	179.7	