


آزمایشگاه کولایف تهران

Control Serum - Normal Levey Jennings Chart

Day-Run	Value	Description
3 - 1	2.988	
3 - 2	2.892	
3 - 3	3.01	
3 - 6	3.12	
3 - 9	2.783	
3 - 10	3.018	
3 - 13	2.91	
5 - 5	3.06	
5 - 6	3.05	
9 - 1	3.051	
10 - 1	2.889	
11 - 1	3.136	
12 - 1	3.091	
13 - 1	2.991	
14 - 1	2.996	
16 - 1	2.998	
17 - 1	3.015	Positive Shift Warning
18 - 1	3.063	Positive Shift Reject
19 - 1	3.063	
19 - 2	2.96	
20 - 1	2.897	
21 - 1	3.052	
23 - 1	3.121	
24 - 1	2.97	
25 - 1	3.075	
26 - 1	2.912	
30 - 1	2.855	
31 - 1	3.062	
6 - 1	2.845	
2 - 1	3.015	
3 - 14	3.112	
4 - 1	2.852	
6 - 2	3.167	

WHO	تفسیر	Free PSA	آزمایش	Architect	دستگاه	6	ماه	1400	سال
Control	multichem	Mean	2.98	SD	0.13	Exp. Date	1401/1/21	Lot. No	33402200
Next Month	Mean	2.99	SD	0.1	CV	3.18			



Day-Run	Value	Description
7 - 1	2.996	
8 - 1	2.966	
9 - 2	3.038	
10 - 2	2.978	
11 - 2	3.047	
13 - 2	2.987	
14 - 2	2.998	
15 - 1	2.941	
16 - 2	2.795	
17 - 2	2.907	
18 - 2	2.793	