

آزمایشگاه فرجاد تهران – آزمایشگاه کولایف

Control Chart

Date-Run	Value
1401/9/2-1	1.46
1401/9/3-1	1.42
1401/9/5-1	1.45
1401/9/6-1	1.46
1401/9/7-1	1.43
1401/9/8-1	1.45
1401/9/9-1	1.46
1401/9/10-1	1.43
1401/9/12-1	1.46
1401/9/13-1	1.42
1401/9/14-1	1.43
1401/9/15-1	1.45
1401/9/16-1	1.46
1401/9/17-1	1.48
1401/9/24-1	1.45
1401/9/26-1	1.46
1401/9/27-1	1.42
1401/9/28-1	1.43
1401/9/29-1	1.44
1401/9/30-1	1.41
1401/10/1-1	1.44
1401/10/3-1	1.46
1401/10/20-1	1.45
1401/10/21-1	1.46
1401/10/22-1	1.43
1401/10/24-1	1.46
1401/10/25-1	1.45
1401/10/26-1	1.44
1401/10/27-1	1.42
1401/10/28-1	1.45
1401/10/29-1	1.47
1401/11/1-1	1.45
1401/11/2-1	1.42
1401/11/3-1	1.44
1401/11/4-1	1.43

Normal					
mindray BS-800M	دستگاه	CR	تست	1401/9/2	تاریخ
1:3.5s	قوانین تفسیر	Control	Randonx	Lot. No	1489UN

Mean	1.44	SD	0.03	Calculated :	Mean	1.44	SD	0.02	CV %	1.18 (8.87)	Sigma	11.48	N	53
------	------	----	------	--------------	------	------	----	------	------	-------------	-------	-------	---	----



Date-Run	Value
1401/11/5-1	1.47
1401/11/6-1	1.41
1401/11/8-1	1.46
1401/11/9-1	1.45
1401/11/10-1	1.43
1401/11/11-1	1.47
1401/11/12-1	1.45
1401/11/13-1	1.47
1401/11/16-1	1.43
1401/11/17-1	1.45
1401/11/18-1	1.43
1401/11/19-1	1.45
1401/11/20-1	1.42
1401/11/23-1	1.43
1401/11/24-1	1.45
1401/11/25-1	1.45
1401/11/26-1	1.43
1401/11/27-1	1.44