

# آزمایشگاه کولایف تهران

## Control Chart

Date-Run	Value
1400/9/1-1	207
1400/9/3-1	206
1400/9/4-1	210
1400/9/6-1	208
1400/9/7-1	205
1400/9/8-1	206
1400/9/9-1	204
1400/9/10-1	207
1400/9/11-1	213
1400/9/13-1	211
1400/9/14-1	212
1400/9/15-1	205
1400/9/16-1	205
1400/9/20-1	211
1400/9/21-1	211
1400/9/22-1	210
1400/9/24-1	203
1400/9/25-1	209
1400/9/27-1	213

High					
mindray BS-800M	دستگاه	Iron	تست	1400/9/1	تاریخ
1:3s	قوانین تفسیر	Control	Randox Hum Asy L3	Lot. No	1173UE

Mean	209	SD	8.85	Calculated :	Mean	208.21	SD	3.15	CV %	1.52 (13.3)	Sigma	19.99	N	19
------	-----	----	------	--------------	------	--------	----	------	------	-------------	-------	-------	---	----

