

آزمایشگاه کولایف تهران

Control Chart

Date-Run	Value
1400/9/4-1	5.66
1400/9/6-1	5.67
1400/9/7-1	5.66
1400/9/8-1	5.7
1400/9/9-1	5.6
1400/9/10-1	5.7
1400/9/11-1	5.66
1400/9/13-1	5.7
1400/9/14-1	5.65
1400/9/16-1	5.67
1400/9/20-2	5.77
1400/9/21-3	5.57
1400/9/22-4	5.59
1400/9/23-5	5.67
1400/9/24-6	5.69
1400/9/25-7	5.77

High					
Easy Lyte Plus	دستگاه	potassium(k)	تست	1400/9/1	تاریخ
1:3s	قوانین تفسیر	Control	Easy Lyte	Lot. No	21054

Mean	5.67	SD	0.1	Calculated :	Mean	5.67	SD	0.05	CV %	0.97 (2.3)	Sigma	5.80	N	16
------	------	----	-----	--------------	------	------	----	------	------	------------	-------	------	---	----

