

آزمایشگاه کولایف تهران

Control Chart

Date-Run	Value
1400/9/4-1	3.9
1400/9/6-1	3.92
1400/9/7-1	3.91
1400/9/8-1	3.84
1400/9/9-1	3.85
1400/9/10-1	3.9
1400/9/11-1	3.94
1400/9/13-1	3.92
1400/9/14-1	3.82
1400/9/16-1	3.91
1400/9/20-1	3.93
1400/9/21-1	3.94
1400/9/22-1	3.9
1400/9/23-1	3.85
1400/9/24-1	3.92
1400/9/25-1	3.92

Normal					
Easy Lyte Plus	دستگاه	potassium(k)	تست	1400/9/1	تاریخ
1:2.5s	قوانین تفسیر	Control	Easy Lyte	Lot. No	21047

Mean	3.9	SD	0.1	Calculated :	Mean	3.90	SD	0.04	CV %	0.96 (2.3)	Sigma	5.31	N	16
------	-----	----	-----	--------------	------	------	----	------	------	------------	-------	------	---	----

