

آزمایشگاه کولایف تهران

Control Chart

Date-Run	Value
1400/9/3-1	154
1400/9/4-1	153.8
1400/9/6-1	154
1400/9/7-1	153.5
1400/9/8-1	154
1400/9/10-1	154.5
1400/9/11-1	154.2
1400/9/13-1	154.1
1400/9/14-1	154.3
1400/9/16-1	154.3
1400/9/20-1	154.4
1400/9/21-1	153.9
1400/9/22-1	153.8
1400/9/23-1	154.2
1400/9/24-1	153.7
1400/9/25-1	155

High					
Easy Lyte Plus	دستگاه	Sodium (Na)	تست	1400/9/1	تاریخ
1:3s	قوانین تفسیر	Control	Easy Lyte	Lot. No	21054

Mean	154	SD	1	Calculated :	Mean	154.11	SD	0.36	CV %	0.23 (0.3)	Sigma	11.82	N	16
------	-----	----	---	--------------	------	--------	----	------	------	------------	-------	-------	---	----

