

# آزمایشگاه پاتوبیولوژی و ژنتیک فرجاد

## Control Serum - Levey Jennings Chart

Normal		High - CV : 7.96	
Mean	1.8	Mean	17.91
SD	0.12	SD	1.43
CV	6.62	CV	7.96
Day-Run	Value	Day-Run	Value
1 - 1	1.82	1 - 1	17.34
2 - 1	1.98	2 - 1	16.9
13 - 1	1.92	13 - 1	20.6
18 - 1	1.95	18 - 1	15.35
19 - 1	1.69	19 - 1	16.3
20 - 1	1.61	20 - 1	17.46
23 - 1	1.798	23 - 1	19.08
24 - 1	1.85	24 - 1	18.42
26 - 1	1.7	26 - 1	19.49
30 - 1	1.73	29 - 1	17.59
31 - 1	1.72	30 - 1	18.1
		31 - 1	18.32

Westgard	تفسیر	TSH	آزمایش	ADVIA Immunoassay	دستگاه	3	ماه	1400	سال
Normal Control	Indox Premium Plus	Mean	1.8	SD	0.13	Exp. Date	1400/1/30	Lot. No	1765EC

